



Drummond Scientific Company

500 PARKWAY, BROOMALL, PENNSYLVANIA 19008-4203
PHONE : 610 – 353 – 0200 FAX 610 – 353 – 6204
www.drummondsci.com

September 6, 2022

Certificate of Conformance & Sterility

Products listed conform to Drummond Scientific Company specifications and drawings, and have passed inspection and testing criteria for sterility. This item has been manufactured in accordance with FDA GMP's as Medical Device Materials of construction have passed USP Class VI-121 C. plastics test for biosafety

<u>Item</u>	<u>Part #</u>	<u>Lot</u>	<u>Exp</u>
Sterile Drummond TC Filter	4-000-051-S	82020	12/01/2023

Method of Sterilization	Gamma Radiation
Hydrophilic Pore Size	. 8 um PES
Hydrophobic Pore Size	. 8 um PTFE
Filter Diameter	24 mm
Housing Material	Polypropylene
Bubble point	>1 bar

Matt Herm
Quality Assurance